




Instrument Expert Original factory packaging www.dorgean.com





# **Precision Pulse Control**

The PCX-7401 offers the capability of providing both pulsed and bias outputs. A modern internal trigger source is capable of two modes of operation: duty cycle and single shot. External trigger is available for additional flexibility.

# Low Inductance Output Cable

Connection to the laser diode is made through an innovative low-inductance stripline cable, designed to preserve the fidelity of high-speed current pulses. The output connector is interlocked so that the PCX-7401 is disabled when the cable is removed.

# **Output Protection**

The PCX-7401 features advanced circuitry to protect both the laser diode and instrument. At turn on, and at any time the output is not enabled, the PCX-7401's output is electronically shorted to ground, ensuring that no current flows through the laser diode. Safety features of the instrument include a separate output enable key switch, an output cable safety interlock, and an external enable control signal.

# **Ease of Setup and Operation**

The PCX-7401 may be operated through the intuitive front panel controls. The color LCD provides immediate visual confirmation of all operating parameters.

# **Store and Recall User Settings**

All system configurations may be stored and recalled in the internal non-volatile memory.


# **Complete System Integration**

Automated applications can utilize RS-232, USB, or Ethernet computer interfaces.


# **Ordering Information**

PCX-7401	Precision Pulsed Current Source
6045-0003	Output Stripline Cable
6045-0097	Laser Output PCBA
PCA-9550	Current/Voltage Monitor Cable
PCA-9410	BNC Shorting Connector
Fach DCV 7401 is	delivered with a Output Stripling Cable Las


Each PCX-7401 is delivered with a Output Stripline Cable, Laser Output PCBA, Current/Voltage Monitor Cable and BNC shorting connector.



3.000 A output with 0.500 A bias







# Laser Output PCBA

On the laser output PCBA above, the current monitor (J1) has a ratio of 125 mV/A, with a 50  $\Omega$  termination.





Instrument Expert **Original factory packaging** www.dorgean.com

# **PCX-7401** Pulsed Precision Current Source — Datasheet



Requires 50 Ω

≤ 2,000,000 Hz

 $50~\Omega$  or  $10~k\Omega$ 

47 Hz to 63 Hz

NEMA C-14

15° C to 40° C

7.8 kg

100 VAC to 120 VAC ± 10% 220 VAC to 240 VAC ± 10%

35 A/115 VAC 70A/230 VAC

10.66 cm x 29.21 cm x 51.06 cm

5 V = Output to load

0 V = No output to load

0 V to 4.5 V

~ 100 ns

100 ns

BNC

~ 130 ns

0 V to 5 V

BNC

# **Pulse Amplitude**

**Output Current Range** Setpoint Resolution Setpoint Accuracy\* **Compliance Voltage** Overshoot Maximum Output Power

# **Bias Amplitude**

**Bias Current Amplitude Bias Current Resolution Bias Current Accuracy** 

# **Output Parameters**

Pulse Width Range **Rise/Fall Time** Polarity

# **Internal Trigger**

Frequency Range	5 Hz to 1.000 MHz	
Frequency Resolution	5 Hz to 995 Hz:	5 Hz
	1 kHz to 49.9 kHz:	100 Hz
	50 kHz to 1 MHz:	1000 Hz
Frequency Accuracy	± (0.01 x setpoint +2) Hz	
$T_{jit(cc)}$ (cycle to cycle jitter)	≤ 25 ns	

0.000 A to 3.000 A

0.000 A to 0.550 A

0.001 A

≤ 15 V

< 2 %

54 W

0.001 A

± 0.001 A

100 ns to DC

< 100 ns

Positive

± 0.001 A

Duty Cycle Range 1 % to 99 % **Duty Cycle Resolution** 0.01 % ± (0.01 x setpoint + 2.5) % Duty Cycle Accuracy

# **Internal Single Shot Trigger**

Pulse Width Range 200 ns to 1.0000 s

Pulse Width Resolution	200 ns to 5,000 ns 6 μs to 1,000 ms	100 ns 1 μs
Pulse Width Accuracy	200 ns to 5,000 ns 6 µs to 50 µs 51 µs to 250 µs 251 µs to 500 µs 501 µs to 2,000 µs 2001 µs to 10,000 µs 10,000 µs to 65,535 µs 65.536 ms to 100 ms 100.001 ms to 1,000 ms	± 5 ns ± 100 ns ± 250 ns ± 2 μs ± 5 μs ± 50 μs ± 250 μs ± 500 μs ± 2,000 μs
Pulsed Bias Output**	Main Pulsewidth 200 ns to 100 μs 100.1 μs to 350 μs 350.1 μs to 1,000 ms	Fixed Bias 2 μs 10 μs 25 μs

# **Trigger Sync Output**

Termination Connector **Output Voltage Levels** Delay (sync to output)

# **External Trigger**

**Frequency Range** Minimum Pulsewidth Delay (external to output) **Termination Impedance** Connector Input Voltage Levels

# **Computer Interface**

RS232, Ethernet, USB

**USB** Driver Support

Windows 8, Windows 7, Windows XP, Linux, and MAC OS X

# General

Power	Requirements	

AC Inrush Current (typical) AC Connector Type

Size (H x W x D) Weight

**Operating Temperature** Cooling

Air cooled Color LCD with touch screen

User interface

# **Notes**

\* Current accuracy +/- 1 mA for output currents with pulse frequency below 100 kHz. Current accuracy between 100 kHz and 1 MHz is: +0 mA / -X mA Where X = ((Output frequency in Hz) \* (current setpoint in A) / 31,000) \*\* In single shot mode bias is a fixed pulsewidth before and after the main pulsewidth, as shown above.

Warranty—One year parts and labor on defects in materials and workmanship.

The PCX-7401 current source meets or exceeds these specifications

All specifications are measured with a low inductance stripline interconnect cable to the laser diode, with less than 4 nH total inductance

Specifications information subject to change without notice.