



# HBM Discharge Simulator PDS-510

## Data Sheet

### HBM Simulator used to test devices in compliance with ANSI/ESDA/JEDEC JS-001

The PDS-510 is a manual two pin tester which means that only two pins at a time can be tested. The user must design a way to connect the PDS-510 to the pins under test. In addition, they must have a way to check for damage to the device either by using a curve tracer or functional tester. HBM data collected using the PDS-510 can be used for data sheet reporting.




**Data Sheet**
**Specifications for the PDS-510 HBM Discharge Simulator**

PDS-510 HBM Discharge Simulator	
Discharge Output	Adjustable up to 4kV
Power Supply Output	Adjustable up to 4kV
Power Supply Accuracy	±2%
Discharge Waveform ANSI/ESD STM11.31	Rise time @ 0Ω: <10 nsec @ 500Ω: <20 nsec Fall time @ 0Ω: 150±15 nsec @ 500Ω: 200±20 nsec Ringing: <15%
Power	<b>Batteries:</b> 4 each AA 2600 mAh low self-discharge, Rechargeable, Nickel-Metal Hydride Battery <b>Power Supply:</b> From 90VAC to 264VAC, 47Hz to 63Hz. UL and TUV 60950-1 approved.
Battery Life	> 100 hours (with average of 50 discharges in 8 hours)
Dimensions	7.5"L x 4.5"W x 2.5"H (19.05cm x 11.43cm x 6.35cm)
Weight	2.0 lbs (907g)

ANSI/ESDA/Jedec JS-001-2014						
Voltage Level (V)	Ipeak for Short, Ips (A)	Ipeak for 500 Ω Ipr (A)	Rise Time for Short, tr (ns)	Rise Time for 500 Ω trr (ns)	Decay Time for Short, Td (ns)	Maximum Ringing Current IR (A)
125 (optional)	0.075 - 0.092 <b>.082</b>	N/A	2.0 - 10 <b>5.0</b>	N/A	130 - 170 <b>153.7</b>	15% of Ips
250	0.15 - 0.18 <b>.169</b>	N/A	2.0 - 10 <b>4.5</b>	N/A	130 - 170 <b>149.7</b>	15% of Ips
500	0.30 - 0.37 <b>.307</b>	N/A	2.0 - 10 <b>5.0</b>	N/A	130 - 170 <b>159.4</b>	15% of Ips
1000	0.60 - 0.73 <b>.654</b>	0.37 - 0.55 <b>0.435</b>	2.0 - 10 <b>4.5</b>	5.0 - 25 <b>10.0</b>	130 - 170 <b>151.3</b>	15% of Ips
2000	1.20 - 1.47 <b>1.302</b>	N/A	2.0 - 10 <b>4.5</b>	N/A	130 - 170 <b>151.0</b>	15% of Ips
4000	2.40 - 2.93 <b>2.72</b>	1.5 - 2.2 <b>1.728</b>	2.0 - 10 <b>4.0</b>	5.0 - 25 <b>10.0</b>	130 - 170 <b>143.3</b>	15% of Ips

The PDS-510 HBM Discharge Simulator meets ANSI/ESDA/Jedec JS-001-2014 Standard. The numbers in red are results from the PDS-510 with a PicoScope® and the Pico software<sup>1</sup>.

<sup>1</sup> Data Graphs available upon request. No 8000 volt discharge available on the PDS-510. PicoScope and Pico Technology are registered trademarks of Pico Technology Ltd.

Phone +1 (630) 238-8883 Fax +1 (630) 238-9717

[www.prostatcorp.com](http://www.prostatcorp.com)

©2019 Prostat Corporation. All rights reserved. Prostat, Prostat Corporation, and the Prostat logo are trademarks or registered trademarks of Prostat Corporation or one of its affiliated companies in the United States and/or other countries. All other trademarks or registered trademarks are the property of their respective owners. Complying with all applicable copyright laws is the responsibility of the user. Prostat reserves the right to change, without notice, product offerings or specifications. Printed in U.S.A. REV1: 4/3/19